



Certificate of Analysis

COA No: CA_RES-0028

Version: 02

ISOLATE II Genomic DNA Kit

For Research Use Only

Storage Conditions:	Room Temperature
Lot number:	IS504-BO40660
Expiry date:	12 months from date of purchase

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II Genomic DNA Spin Columns	Nucleic acid recovery >90%	Passed	SC-416J
Collection tubes	n/a	n/a	CT-416J
Lysis Buffer GL	pH = 7.9-8.1, density	Passed	GL-416J
Buffer G1	pH = 3.7-3.9, density	Passed	G1-416J
Buffer G2	density	Passed	G2-416J
Wash Buffer GW1	pH = 7.2-7.6, density	Passed	GW1-416J
Wash Buffer GW2	pH = 7.5-7.9, density, sterility	Passed	GW2-416J
Elution Buffer G	pH = 8.3-8.7, density, sterility	Passed	EBG-416J
Proteinase K	activity	Passed	PK-416J
Proteinase K Buffer	pH = 3.5-6.0, density	Passed	PKB-416J

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

USA

Germany

France

Australia

Singapore

info.uk@bioline.com

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

info.us@bioline.com

Tel: +1 508 880 8990
Fax: +1 508 880 8993

info.de@bioline.com

Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

info.fr@bioline.com

Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

info.aust@bioline.com

Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

info.sg@bioline.com

Tel: +65 6774 7196
Fax: +65 6774 6441