



A Meridian Life Science® Company

PO Box 122  
Alexandria, NSW 1435  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

## CERTIFICATE OF ANALYSIS

Catalog #                    BIO-38041  
Product Name:                RNA Extraction Control 670  
Product Description:        500 reactions RNA Extraction Control 670 Kit  
Batch #:                      CT564-BO08560  
Expiry date:                 NOVEMBER 2017

### QC Methods: Functional test

RNA Extraction Control 670 kit was functionally tested based on the recommended procedures listed on the product package insert. The Internal Control RNA was analyzed on the LightCycler® 480 II Real-Time PCR System. RT-PCR amplification was detected on excitation 618nm and emission 660nm configuration.

### Specifications:

Analysis	Results
RNA functional test	Pass

### Result


Product meets all specification

This product is for research use only, not intended for diagnostic use

Released by: \_\_\_\_\_

Title: \_\_\_\_\_

Date: \_\_\_\_\_

  
QA/RA Senior Scientist

11/01/2017

**Trademarks:** LightCycler® is trademark of Roche.