



Certificate of Analysis

COA No: CA_RES-0019

Version: 02

ISOLATE II Plasmid Mini Kit

For Research Use Only

Storage Conditions:	Room Temperature
Lot number:	IS494-BO49820
Expiry date:	12 months from date of purchase

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II Plant DNA Spin Columns	Nucleic acid recovery >90%	Passed	SPI002242
Collection tubes	n/a	n/a	COL001615
Resuspension Buffer P1	pH = 7.9-8.1, density	Passed	73864
Lysis Buffer P2	density	Passed	73869
Neutralization Buffer P3	pH = 4.1-4.4, density	Passed	73863
Wash Buffer PAW1	pH = 7.2-7.6, density	Passed	73877
Wash Buffer PAW2	pH = 7.5-7.9, density, sterility	Passed	73855
Elution Buffer P	pH = 8.3-8.7, density, sterility	Passed	73860
RNase A	no DNase activity detectable in connection with our kit	Passed	R1730S

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

[Info.sg@bioline.com](mailto:info.sg@bioline.com)
Tel: +65 6774 7196
Fax +65 6774 6441